



Certificate of Analysis

1.02149.0001 Contact TVC (Total Viable Counts) Envirocheck®
Batch 1916718

	Spec. Values	Batch Values
Microbial contamination	no growth	passes test
pH-value/side one (25 °C)	6.7 - 7.3	7.2
pH-value/side two (25 °C)	6.7 - 7.3	7.1

	Spec. Values	Batch Values
Growth/side one (Escherichia coli ATCC 8739 (WDCM 00012))	good	good
Growth/side one (Staphylococcus aureus ATCC 6538 (WDCM 00032))	fair/good	fair/good
Growth/side one (Candida albicans ATCC 10231 (WDCM 00054))	none/poor	none/poor
Growth/side one (Bacillus subtilis ATCC 6633 (WDCM 00003))	good	good
Growth/side one (Pseudomonas aeruginosa ATCC 9027 (WDCM 00026))	good	good
Growth/side one (Aspergillus brasiliensis (formerly A. niger) ATCC 16404 (WDCM 00053))	good	good
Growth/side two (Escherichia coli ATCC 8739 (WDCM 00012))	good	good
Growth/side two (Staphylococcus aureus ATCC 6538 (WDCM 00032))	good	good
Growth/side two (Candida albicans ATCC 10231 (WDCM 00054))	good	good
Growth/side two (Bacillus subtilis ATCC 6633 (WDCM 00003))	good	good
Growth/side two (Pseudomonas aeruginosa ATCC 9027 (WDCM 00026))	good	good
Growth/side two (Aspergillus brasiliensis (formerly A. niger) ATCC 16404 (WDCM 00053))	good	good

Incubation: 48 hrs.; 35 °C; aerobic; Aspergillus up to 72 hrs.

Date of release (DD.MM.YYYY) 07.09.2023
Expiry date (DD.MM.YYYY) 11.04.2024

Dr. Lukas Mechler
Responsible laboratory manager quality control

This document has been produced electronically and is valid without a signature.