



Certificate of Analysis

1.01299.0001 Enterococci 100 Readycult®
Batch K95851599

	Spec. Values	Batch Values
Appearance (clearness)	clear	clear
Appearance (color)	yellowish to yellowish-brown	yellow
pH-value (25 °C)	7.3 - 7.7	7.4
Sterility test (48 hrs., 35 °C)	passes test	passes test

	Spec. Values	Batch Values
Inoculum on reference medium (Enterococcus faecalis ATCC 19433 (WDCM 00009))	10 - 100	97
Inoculum on reference medium (Enterococcus durans ATCC 19432)	10 - 100	17
Inoculum on reference medium (Staphylococcus aureus ATCC 25923 (WDCM 00034))	≥ 1000	≥ 1000
Growth (Enterococcus faecalis ATCC 19433 (WDCM 00009))	fair to good	good
Growth (Enterococcus durans ATCC 19432)	fair to good	good
Growth (Staphylococcus aureus ATCC 25923 (WDCM 00034))	fair to good	moderately
Color change to green-blue (Enterococcus faecalis ATCC 19433 (WDCM 00009))	+	+
Color change to green-blue (Enterococcus durans ATCC 19432)	+	+
Color change to green-blue (Staphylococcus aureus ATCC 25923 (WDCM 00034))	-	-

Incubation: 24 - 48 hrs.; 35 °C; aerobic.

Date of release (DD.MM.YYYY) 24.08.2023

Expiry date (DD.MM.YYYY) 22.04.2026

Dr. Lukas Mechler

Responsible laboratory manager quality control

This document has been produced electronically and is valid without a signature.