



# Certificate of Analysis

1.02137.0001 Contact E (Enterobacteriaceae) Envirocheck®  
Batch 1916719

	Spec. Values	Batch Values
Microbial contamination	no growth	passes test
pH-value/side one (25 °C)	6.7 - 7.3	7.1
pH-value/side two (25 °C)	6.9 - 7.6	7.1

	Spec. Values	Batch Values
Growth/side one (Escherichia coli ATCC 8739 (WDCM 00012))	good	good
Growth/side one (Staphylococcus aureus ATCC 6538 (WDCM 00032))	good	good
Growth/side one (Salmonella typhimurium ATCC 14028 (WDCM 00031))	good	good
Growth/side one (Shigella sonnei ATCC 25931)	good	good
Growth/side two (Escherichia coli ATCC 8739 (WDCM 00012))	fair/good	fair/good
Growth/side two (Staphylococcus aureus ATCC 6538 (WDCM 00032))	none	none
Growth/side two (Salmonella typhimurium ATCC 14028 (WDCM 00031))	fair/good	fair/good
Growth/side two (Shigella sonnei ATCC 25931)	fair/good	fair/good

Incubation: 48 hrs.; 35 °C; aerobic.

Date of release (DD.MM.YYYY) 13.09.2023  
Expiry date (DD.MM.YYYY) 11.02.2024

Dr. Lukas Mechler  
Responsible laboratory manager quality control

This document has been produced electronically and is valid without a signature.