



Certificate of Analysis

1.02136.0001 Contact C (total coliforms/E.coli) Envirocheck®
Batch 1916768

	Spec. Values	Batch Values
Microbial contamination	no growth	passes test
pH-value/side one (25 °C)	6.7 - 7.3	7.1
pH-value/side two (25 °C)	6.5 - 7.1	6.9

	Spec. Values	Batch Values
Growth/side one (Escherichia coli ATCC 11775 (WDCM 00090))	good	good
Growth/side one (Escherichia coli O157:H7 ATCC 35150)	good	good
Growth/side one (Citrobacter freundii ATCC 8090)	good	good
Growth/side one (Salmonella enteritidis ATCC 13076 (WDCM 00030))	good	good
Growth/side two (Escherichia coli ATCC 11775 (WDCM 00090))	good	good
Growth/side two (Escherichia coli O157:H7 ATCC 35150)	fair/good	fair/good
Growth/side two (Citrobacter freundii ATCC 8090)	good	good
Growth/side two (Salmonella enteritidis ATCC 13076 (WDCM 00030))	good	good
Colonies (Escherichia coli ATCC 11775 (WDCM 00090))	dark blue to violet	dark blue to violet
Colonies (Escherichia coli O157:H7 ATCC 35150)	salmon to red	salmon to red
Colonies (Citrobacter freundii ATCC 8090)	salmon to red	salmon to red
Colonies (Salmonella enteritidis ATCC 13076 (WDCM 00030))	colourless	colourless

Incubation: 48 hrs.; 35 °C; aerobic.

Date of release (DD.MM.YYYY) 13.09.2023
Expiry date (DD.MM.YYYY) 17.04.2024

Dr. Lukas Mechler
Responsible laboratory manager quality control

This document has been produced electronically and is valid without a signature.